

01/14/11 Recall

Urgent Product Recall Notification

We are requesting the return of lot # 10203-012 manufactured **Triad** by as follows:

Level of Recall:	Voluntarily Retail level recall
Product:	Lubricating Jelly Sterile 4oz
Reason:	Manufacture Triad is recalling due to concerns expressed by the FDA regarding validation of the gamma radiation sterilization cycles for these products. Pharmedix is initiating this recall because use of inadequately sterilized product might result in patient infection.
Lot #s	10203-012
Mfg	Triad
Mfgs. NDC #	50730-8500-0

This voluntary recall is extended to the retail level and is being made with the knowledge of the Food and Drug Administration.

Our records show that product with the referenced Lot #s has been shipped to your clinic. Please examine your inventory to verify if you have product from these Lots remaining in stock. If so, **please discontinue dispensing it and promptly call Pharmedix at 1-800-486-1811, Ext. 11** to obtain the approved paperwork to return all concerned product to us for credit to your account.

Please check the appropriate box below, sign where indicated **and fax this form back to Pharmedix at 1-800-783-2038 even if you do not have any product from the indicated Lot # remaining in stock.**

Sincerely,

Tonya Tarneem

Client Services

800-486-1811x11

Clinic Code	Order Code	PMX Lot #	Rx #	Quantity

Check appropriate box:

Received affected Lot Number(s) – Do not have any stock on hand.

Received affected Lot Number(s) – Have ____ units remaining in stock.

Person verifying stock status (print): _____

Date: _____

Signature: _____